



SAFETY AND BUILDINGS DIVISION  
Plumbing Product Review  
P.O. Box 2658  
Madison, Wisconsin 53701-2658  
TTY: Contact Through Relay

Jim Doyle, Governor  
Richard J. Leinenkugel, Secretary

December 30, 2008

A-DEC INC.  
ENGINEERING  
TOM LOUISELL  
2601 CRESTVIEW DRIVE  
PO BOX 111  
NEWBERG OR 97132

Re: Description: HEALTH CARE PLUMBING APPLIANCE  
Manufacturer: A-DEC INC.  
Product Name: DENTAL DELIVERY SYSTEM  
Model Number(s): DECADE DOCTOR'S CART: 2542, 2545 AND 22512; DOCTOR'S IC MICRO-CART: 2503  
AND 2523; CASCADE CART: 2615 AND 2671; AND DECADE DUO CART 2541, 2544  
AND 22511  
Product File No: 20080371

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters Comm 82 through 84, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of May 2014.

This approval is contingent upon compliance with the following stipulation(s):

- A reduced pressure principle backflow preventer (ASSE 1013 or CAN/CSA B64.4) or air-gap must be installed on the water supply pipes serving this product.

This approval supersedes the approval issued on May 4, 2004 under product file number 20040194.

This approval letter shall be incorporated with your previously approved plans and/or specifications approved under product file number 20040194.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Jerry Thompson  
Plumbing Product Reviewer  
Phone: 608-266-6742  
Fax: 608-267-9566  
E-mail: Jerry.Thompson@wi.gov